

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052019\
 Data File : VX009725.D
 Acq On : 20 May 2019 14:52
 Operator : JC/SP
 Sample : K2952-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB202-GW-698-700

Quant Time: May 21 08:19:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	176590	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	280910	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	248059	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	103850	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	116816	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
35) Dibromofluoromethane	5.50	113	85820	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
50) Toluene-d8	8.71	98	354674	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
62) 4-Bromofluorobenzene	11.13	95	117312	45.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.56%	
Target Compounds						
16) Acetone	2.44	43	13116	10.133	ug/l	93
52) Toluene	8.79	92	6890	1.403	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052019\
Data File : VX009725.D
Acq On : 20 May 2019 14:52
Operator : JC/SP
Sample : K2952-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB202-GW-698-700

Quant Time: May 21 08:19:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

