

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052019\  
 Data File : VX009734.D  
 Acq On : 20 May 2019 18:23  
 Operator : JC/SP  
 Sample : K2910-06  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 S81-0102

Quant Time: May 21 08:35:46 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X042919W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 09 07:31:15 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units  | Dev(Min) |
|-----------------------------|-------|------|----------|--------|--------|----------|
| 1) Pentafluorobenzene       | 5.67  | 168  | 167975   | 50.00  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.86  | 114  | 269667   | 50.00  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.11 | 117  | 236343   | 50.00  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.08 | 152  | 95040    | 50.00  | ug/l   | 0.00     |
| System Monitoring Compounds |       |      |          |        |        |          |
| 33) 1,2-Dichloroethane-d4   | 6.06  | 65   | 113108   | 49.33  | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 98.66% |          |
| 35) Dibromofluoromethane    | 5.50  | 113  | 83046    | 48.99  | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 97.98% |          |
| 50) Toluene-d8              | 8.71  | 98   | 339294   | 49.57  | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 99.14% |          |
| 62) 4-Bromofluorobenzene    | 11.13 | 95   | 108875   | 43.78  | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 87.56% |          |
| Target Compounds            |       |      |          |        |        |          |
| 16) Acetone                 | 2.44  | 43   | 3004     | 2.440  | ug/l   | 94       |
| 27) cis-1,2-Dichloroethene  | 4.59  | 96   | 25369    | 12.434 | ug/l   | 87       |
| 40) Benzene                 | 6.14  | 78   | 31710    | 4.089  | ug/l   | 98       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052019\  
Data File : VX009734.D  
Acq On : 20 May 2019 18:23  
Operator : JC/SP  
Sample : K2910-06  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
S81-0102

Quant Time: May 21 08:35:46 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X042919W.M  
Quant Title : SW846 8260  
QLast Update : Thu May 09 07:31:15 2019  
Response via : Initial Calibration

