

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052019\
 Data File : VX009735.D
 Acq On : 20 May 2019 18:47
 Operator : JC/SP
 Sample : K2910-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S81-0103

Quant Time: May 21 08:38:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	177905	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	282328	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	246788	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	100905	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	118586	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
35) Dibromofluoromethane	5.50	113	86159	48.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.10%	
50) Toluene-d8	8.71	98	354997	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
62) 4-Bromofluorobenzene	11.13	95	115545	44.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.76%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.41	62	6395	2.930	ug/l	94
11) Tert butyl alcohol	3.03	59	4051	7.206	ug/l	96
16) Acetone	2.45	43	7481	5.737	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	14543	6.730	ug/l	89
31) Cyclohexane	5.59	56	4039	1.125	ug/l #	86
39) Methylcyclohexane	7.46	83	1412	0.438	ug/l #	84
40) Benzene	6.15	78	11360	1.399	ug/l #	87
85) sec-Butylbenzene	11.94	105	2480	0.348	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052019\
Data File : VX009735.D
Acq On : 20 May 2019 18:47
Operator : JC/SP
Sample : K2910-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S81-0103

Quant Time: May 21 08:38:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

