

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052019\
 Data File : VX009737.D
 Acq On : 20 May 2019 19:34
 Operator : JC/SP
 Sample : K2910-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S73-0235

Manual Integrations
 APPROVED

MMDadoda
 5/21/2019 9:13:52 AM

Quant Time: May 21 08:41:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	176004	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	280881	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	251627	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	110924	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	118442	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
35) Dibromofluoromethane	5.50	113	86584	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
50) Toluene-d8	8.71	98	356296	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
62) 4-Bromofluorobenzene	11.13	95	119764	46.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.48%	
Target Compounds						
4) Vinyl Chloride	1.41	62	37385	17.314	ug/l	99
16) Acetone	2.45	43	5258	4.076	ug/l	97
21) trans-1,2-Dichloroethene	3.17	96	1436	0.802	ug/l #	79
27) cis-1,2-Dichloroethene	4.60	96	22159	10.366	ug/l	89
44) Trichloroethene	7.22	130	1139	0.583	ug/l	87
85) sec-Butylbenzene	11.94	105	9874	1.261	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052019\
Data File : VX009737.D
Acq On : 20 May 2019 19:34
Operator : JC/SP
Sample : K2910-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleID :
S73-0235

Manual Integrations
APPROVED

MMDadoda
5/21/2019 9:13:52 AM

Quant Time: May 21 08:41:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

