

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052019\
 Data File : VX009739.D
 Acq On : 20 May 2019 20:21
 Operator : JC/SP
 Sample : K2910-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S73-ER-2019-1

Quant Time: May 21 08:45:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	172367	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	274479	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	243465	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	103157	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	116687	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
35) Dibromofluoromethane	5.50	113	84867	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
50) Toluene-d8	8.71	98	346531	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
62) 4-Bromofluorobenzene	11.13	95	114703	45.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.62%	
Target Compounds						
16) Acetone	2.45	43	5381	4.259	ug/l	92
20) Methylene Chloride	2.85	84	15905	7.775	ug/l	98
29) Tetrahydrofuran	5.13	42	58958	46.665	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052019\
Data File : VX009739.D
Acq On : 20 May 2019 20:21
Operator : JC/SP
Sample : K2910-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S73-ER-2019-1

Quant Time: May 21 08:45:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

