

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052019\
 Data File : VX009732.D
 Acq On : 20 May 2019 17:37
 Operator : JC/SP
 Sample : K2952-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB202-GW-678-680

Quant Time: May 21 08:32:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	171038	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	271896	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	241774	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	104871	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	115576	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
35) Dibromofluoromethane	5.50	113	84215	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
50) Toluene-d8	8.71	98	343459	49.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.54%	
62) 4-Bromofluorobenzene	11.13	95	115656	46.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.26%	
Target Compounds						
16) Acetone	2.45	43	6801	5.425	ug/l	Qvalue 90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052019\
Data File : VX009732.D
Acq On : 20 May 2019 17:37
Operator : JC/SP
Sample : K2952-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB202-GW-678-680

Quant Time: May 21 08:32:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.00 min

Lab File: VX009732.D

Acq: 20 May 2019 17:37

Instrument :

MSVOA_X

ClientSampleId :


BP-VPB202-GW-678-680

Tot Ion:168 Resp: 171038

Ion Ratio Lower Upper




168 100

99 53.0 42.3 63.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#16

Acetone

Concen: 5.425 ug/l

RT: 2.45 min Scan# 212

Delta R.T. 0.00 min

Lab File: VX009732.D


Acq: 20 May 2019 17:37

Tgt Ion: 43 Resp: 6801

Ion Ratio Lower Upper



43 100


58 36.8 24.9 37.3



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0









#33
 1,2-Dichloroethane-d4
 Concen: 49.500 ug/l
 RT: 6.06 min Scan# 805
 Delta R.T. 0.00 min
 Lab File: VX009732.D
 Acq: 20 May 2019 17:37




Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB202-GW-678-680


Tot Ion: 65 Resp: 115576
 Ion Ratio Lower Upper
 65 100
 67 53.8 0.0 106.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. 0.00 min
 Lab File: VX009732.D
 Acq: 20 May 2019 17:37

Tgt Ion: 114 Resp: 271896
 Ion Ratio Lower Upper
 114 100
 63 22.7 0.0 45.6
 88 17.3 0.0 33.2









#35
 Dibromofluoromethane
 Concen: 49.270 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. 0.01 min
 Lab File: VX009732.D
 Acq: 20 May 2019 17:37




Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB202-GW-678-680


Tot Ion:113 Resp: 84215
 Ion Ratio Lower Upper
 113 100
 111 102.9 82.8 124.2
 192 21.7 17.0 25.6



#50
 Toluene-d8
 Concen: 49.772 ug/l
 RT: 8.71 min Scan# 1240
 Delta R.T. 0.00 min
 Lab File: VX009732.D
 Acq: 20 May 2019 17:37

Tgt Ion: 98 Resp: 343459
 Ion Ratio Lower Upper
 98 100
 100 64.0 51.2 76.8





#62

4-Bromofluorobenzene

Concen: 46.125 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX009732.D

Acq: 20 May 2019 17:37

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB202-GW-678-680


Tot Ion: 95 Resp: 115656

Ion Ratio Lower Upper

95 100

174 81.2 0.0 161.6




176 78.7 0.0 159.2



Abundance Ion 94.90 (94.60 to 95.60): VX009732.D (-1588) (-)

Ion 173.80 (173.50 to 174.50): VX009732.D (-1588) (-)

Ion 175.80 (175.50 to 176.50): VX009732.D (-1588) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.11 min Scan# 1469

Delta R.T. 0.00 min

Lab File: VX009732.D

Acq: 20 May 2019 17:37


Tgt Ion: 117 Resp: 241774

Ion Ratio Lower Upper

117 100

82 60.4 48.8 73.2



119 31.8 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): VX009732.D (-1420) (-)

Ion 82.00 (81.70 to 82.70): VX009732.D (-1420) (-)

Ion 118.90 (118.60 to 119.60): VX009732.D (-1420) (-)






#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 12.08 min Scan# 1792
 Delta R.T. 0.00 min
 Lab File: VX009732.D
 Acq: 20 May 2019 17:37

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB202-GW-678-680

Tot Ion	Ratio	Lower	Upper
152	100		
115	58.0	41.2	123.6
150	156.5	0.0	346.4



Abundance Ion 151.90 (151.60 to 152.60): V
 Ion 114.90 (114.60 to 115.60): V
 Ion 149.90 (149.60 to 150.60): V

