

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052020\
 Data File : VX016356.D
 Acq On : 20 May 2020 13:44
 Operator : JC/SP
 Sample : L2663-05 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1378

Quant Time: May 21 04:34:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	284125	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	480189	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	471538	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	233452	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	169619	45.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.74%	
35) Dibromofluoromethane	5.46	113	140540	46.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.40%	
50) Toluene-d8	8.70	98	588934	50.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.48%	
62) 4-Bromofluorobenzene	11.12	95	234743	52.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.30%	
Target Compounds						
16) Acetone	2.42	43	27184	16.295	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052020\
Data File : VX016356.D
Acq On : 20 May 2020 13:44
Operator : JC/SP
Sample : L2663-05 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1378

Quant Time: May 21 04:34:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

