

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052020\
 Data File : VX016359.D
 Acq On : 20 May 2020 14:51
 Operator : JC/SP
 Sample : L2663-08DL 4X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1366DL

Quant Time: May 21 04:51:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	247790	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	413045	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	405088	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	205407	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	147044	45.10	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	90.20%	
35) Dibromofluoromethane	5.47	113	121199	46.32	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	92.64%	
50) Toluene-d8	8.70	98	505545	50.14	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	100.28%	
62) 4-Bromofluorobenzene	11.12	95	202929	52.91	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	105.82%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.01	59	6469	7.846	ug/l	97
16) Acetone	2.42	43	255872	175.872	ug/l	93
25) 2-Butanone	4.64	43	58077	24.773	ug/l	98
51) 4-Methyl-2-Pentanone	8.62	43	8948	1.945	ug/l	97
59) 2-Hexanone	9.48	43	11710	3.241	ug/l	95
86) p-Isopropyltoluene	12.05	119	6041	0.443	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052020\
Data File : VX016359.D
Acq On : 20 May 2020 14:51
Operator : JC/SP
Sample : L2663-08DL 4X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1366DL

Quant Time: May 21 04:51:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

