

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052020\
 Data File : VX016362.D
 Acq On : 20 May 2020 16:00
 Operator : JC/SP
 Sample : L2708-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0691

Quant Time: May 21 04:59:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	249843	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	414723	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	411176	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	203981	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	146478	44.55	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	89.10%	
35) Dibromofluoromethane	5.47	113	120573	45.89	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	91.78%	
50) Toluene-d8	8.70	98	513870	50.76	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.52%	
62) 4-Bromofluorobenzene	11.12	95	207779	53.96	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	107.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.18	85	2137	0.808	ug/l	93
4) Vinyl Chloride	1.39	62	2411	0.736	ug/l #	86
16) Acetone	2.42	43	2611	1.780	ug/l #	86
21) trans-1,2-Dichloroethene	3.14	96	3074	1.062	ug/l	91
25) 2-Butanone	4.64	43	5749	2.432	ug/l	92
27) cis-1,2-Dichloroethene	4.57	96	52496	16.591	ug/l	86
44) Trichloroethene	7.19	130	16690	3.933	ug/l	97
59) 2-Hexanone	9.48	43	9316	2.568	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052020\
Data File : VX016362.D
Acq On : 20 May 2020 16:00
Operator : JC/SP
Sample : L2708-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0691

Quant Time: May 21 04:59:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

