

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052020\
 Data File : VX016363.D
 Acq On : 20 May 2020 16:22
 Operator : JC/SP
 Sample : L2708-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0692

Quant Time: May 21 05:02:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	250835	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	415854	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	402713	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	200047	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	146886	44.50	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	89.00%	
35) Dibromofluoromethane	5.46	113	122422	46.47	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	92.94%	
50) Toluene-d8	8.70	98	513437	50.58	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.16%	
62) 4-Bromofluorobenzene	11.12	95	201892	52.29	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	104.58%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.18	85	2785	1.049	ug/l	98
4) Vinyl Chloride	1.39	62	3012	0.916	ug/l	93
16) Acetone	2.42	43	2777	1.886	ug/l	96
17) Carbon Disulfide	2.56	76	4241	0.336	ug/l	98
21) trans-1,2-Dichloroethene	3.15	96	36535	12.574	ug/l	94
27) cis-1,2-Dichloroethene	4.57	96	2634	0.829	ug/l	79
44) Trichloroethene	7.19	130	1443	0.339	ug/l	76
52) Toluene	8.77	92	4407	0.590	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052020\
Data File : VX016363.D
Acq On : 20 May 2020 16:22
Operator : JC/SP
Sample : L2708-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0692

Quant Time: May 21 05:02:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

