

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
 Data File : VX028843.D
 Acq On : 20 May 2022 11:42
 Operator : JC/MD
 Sample : VX0520WBL01
 Misc : 5.00mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0520WBL01

Quant Time: May 20 23:26:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	282652	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	431925	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	384824	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	186830	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	198696	51.283	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.560%
35) Dibromofluoromethane	5.385	113	158344	55.954	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	111.900%
50) Toluene-d8	8.647	98	508757	47.684	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.360%
62) 4-Bromofluorobenzene	11.079	95	210322	52.778	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
Data File : VX028843.D
Acq On : 20 May 2022 11:42
Operator : JC/MD
Sample : VX0520WBL01
Misc : 5.00mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0520WBL01

Quant Time: May 20 23:26:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

