

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
 Data File : VX028844.D
 Acq On : 20 May 2022 12:06
 Operator : JC/MD
 Sample : VX0520MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0520MBL01

Quant Time: May 20 23:26:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	268847	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	434748	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	413886	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	240082	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	199611	54.165	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	108.320%
35) Dibromofluoromethane	5.385	113	146955	51.592	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.180%
50) Toluene-d8	8.653	98	521424	48.554	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.100%
62) 4-Bromofluorobenzene	11.079	95	256290	63.895	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	127.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
Data File : VX028844.D
Acq On : 20 May 2022 12:06
Operator : JC/MD
Sample : VX0520MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0520MBL01

Quant Time: May 20 23:26:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

