

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
 Data File : VX028846.D
 Acq On : 20 May 2022 12:57
 Operator : JC/MD
 Sample : N2941-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 12BR-20220518

Quant Time: May 20 23:27:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	242420	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	420710	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	371562	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	186812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	188918	56.851	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	113.700%
35) Dibromofluoromethane	5.385	113	135688	49.226	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.460%
50) Toluene-d8	8.653	98	499056	48.022	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.040%
62) 4-Bromofluorobenzene	11.079	95	168214	43.336	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.680%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	4725	3.313	ug/l #	86
27) cis-1,2-Dichloroethene	4.501	96	2578	0.781	ug/l	91
44) Trichloroethene	7.129	130	1040	0.322	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
Data File : VX028846.D
Acq On : 20 May 2022 12:57
Operator : JC/MD
Sample : N2941-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
12BR-20220518

Quant Time: May 20 23:27:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

