

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
 Data File : VX028849.D
 Acq On : 20 May 2022 14:36
 Operator : JC/MD
 Sample : N2943-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14

Quant Time: May 20 23:28:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	240150	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	425654	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	377358	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	161112	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	184035	55.905	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.820%
35) Dibromofluoromethane	5.391	113	150086	53.817	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.640%
50) Toluene-d8	8.653	98	551473	52.449	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.900%
62) 4-Bromofluorobenzene	11.079	95	221992	56.527	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	113.060%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	4042	2.861	ug/l	97
17) Carbon Disulfide	2.520	76	3702	0.575	ug/l	# 91
19) Methyl tert-butyl Ether	3.117	73	41233	4.675	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
Data File : VX028849.D
Acq On : 20 May 2022 14:36
Operator : JC/MD
Sample : N2943-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14

Quant Time: May 20 23:28:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

