

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
 Data File : VX028850.D
 Acq On : 20 May 2022 15:00
 Operator : JC/MD
 Sample : N2800-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: May 20 23:29:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	246000	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	431302	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	481040	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	162950	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	166983	49.519	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.040%
35) Dibromofluoromethane	5.391	113	138744	49.099	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.200%
50) Toluene-d8	8.653	98	638543	59.935	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	119.880%
62) 4-Bromofluorobenzene	11.079	95	173719	43.656	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.320%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3563	2.462	ug/l #	85
20) Methylene Chloride	2.788	84	916	0.302	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
Data File : VX028850.D
Acq On : 20 May 2022 15:00
Operator : JC/MD
Sample : N2800-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: May 20 23:29:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

