

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
 Data File : VX028852.D
 Acq On : 20 May 2022 15:46
 Operator : JC/MD
 Sample : N2943-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-41

Quant Time: May 20 23:29:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	232172	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	409430	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	444012	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	175240	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	160316	50.374	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.740%	
35) Dibromofluoromethane	5.391	113	131936	49.183	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 98.360%	
50) Toluene-d8	8.653	98	489176	48.368	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.740%	
62) 4-Bromofluorobenzene	11.079	95	205587	54.424	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 108.840%	
Target Compounds						
						Qvalue
16) Acetone	2.374	43	4353	3.187	ug/l	89
20) Methylene Chloride	2.794	84	1095	0.382	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
Data File : VX028852.D
Acq On : 20 May 2022 15:46
Operator : JC/MD
Sample : N2943-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-41

Quant Time: May 20 23:29:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

