

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
 Data File : VX028864.D
 Acq On : 20 May 2022 20:26
 Operator : JC/MD
 Sample : N2958-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT172S1-20220516

Quant Time: May 20 23:34:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	248268	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	437567	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	385125	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	162400	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	171355	50.35	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.70%
35) Dibromofluoromethane	5.391	113	139638	48.71	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.42%
50) Toluene-d8	8.653	98	510995	47.28	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.56%
62) 4-Bromofluorobenzene	11.085	95	174469	43.22	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.44%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	1359	0.49	ug/l	95
16) Acetone	2.380	43	1818	1.24	ug/l	97
44) Trichloroethene	7.129	130	53065	15.81	ug/l	96
64) Tetrachloroethene	9.281	164	1342	0.46	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
Data File : VX028864.D
Acq On : 20 May 2022 20:26
Operator : JC/MD
Sample : N2958-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT172S1-20220516

Quant Time: May 20 23:34:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

