

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
 Data File : VX028877.D
 Acq On : 21 May 2022 02:38
 Operator : JC/MD
 Sample : N2970-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PDB-BLANK-20220519

Quant Time: May 21 06:08:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) Pentafluorobenzene		5.556	168	317621	50.000	ug/l	0.00
34) 1,4-Difluorobenzene		6.763	114	554231	50.000	ug/l	0.00
63) Chlorobenzene-d5		10.055	117	489357	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4		12.024	152	207514	50.000	ug/l	0.00
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4		5.958	65	225915	51.889	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.780%	
35) Dibromofluoromethane		5.391	113	177884	48.987	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.980%	
50) Toluene-d8		8.653	98	652380	47.652	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.300%	
62) 4-Bromofluorobenzene		11.079	95	230331	45.044	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.080%	
Target Compounds							
						Qvalue	
4) Vinyl Chloride		1.374	62	1629	0.386	ug/l	97
16) Acetone		2.380	43	6024	3.224	ug/l	88
20) Methylene Chloride		2.794	84	1175	0.300	ug/l	96
23) Vinyl Acetate		3.727	43	15345	1.580	ug/l #	91
27) cis-1,2-Dichloroethene		4.501	96	10340	2.391	ug/l	86
37) Ethyl Acetate		4.715	43	1051312	184.549	ug/l	99
44) Trichloroethene		7.129	130	11342	2.668	ug/l	99
48) Methyl methacrylate		7.696	41	5926	1.376	ug/l	86
52) Toluene		8.720	92	5283	0.524	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
Data File : VX028877.D
Acq On : 21 May 2022 02:38
Operator : JC/MD
Sample : N2970-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PDB-BLANK-20220519

Quant Time: May 21 06:08:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

