

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
 Data File : VX028878.D
 Acq On : 21 May 2022 03:01
 Operator : JC/MD
 Sample : N2970-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 02BR-20220519

Quant Time: May 21 06:08:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	316687	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	559186	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	503735	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	218023	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	228163	52.559	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 105.120%	
35) Dibromofluoromethane	5.391	113	179501	48.994	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.980%	
50) Toluene-d8	8.653	98	662069	47.931	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 95.860%	
62) 4-Bromofluorobenzene	11.079	95	239977	46.514	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 93.020%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	5575	1.325	ug/l	# 74
16) Acetone	2.380	43	3943	2.116	ug/l	# 82
17) Carbon Disulfide	2.520	76	6011	0.707	ug/l	# 87
27) cis-1,2-Dichloroethene	4.495	96	111299	25.812	ug/l	89
44) Trichloroethene	7.129	130	262020	61.089	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052022\
Data File : VX028878.D
Acq On : 21 May 2022 03:01
Operator : JC/MD
Sample : N2970-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
02BR-20220519

Quant Time: May 21 06:08:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

