

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052120\
 Data File : VX016381.D
 Acq On : 21 May 2020 14:18
 Operator : JC/SP
 Sample : L2730-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z2-TB2-200520

Quant Time: May 22 09:20:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	231658	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	385779	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	379137	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	188035	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	132401	43.43	ug/l	0.00
Spiked Amount						
			Recovery	=	86.86%	
35) Dibromofluoromethane	5.47	113	109737	44.90	ug/l	0.00
Spiked Amount						
			Recovery	=	89.80%	
50) Toluene-d8	8.70	98	471489	50.07	ug/l	0.00
Spiked Amount						
			Recovery	=	100.14%	
62) 4-Bromofluorobenzene	11.12	95	188251	52.56	ug/l	0.00
Spiked Amount						
			Recovery	=	105.12%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052120\
Data File : VX016381.D
Acq On : 21 May 2020 14:18
Operator : JC/SP
Sample : L2730-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z2-TB2-200520

Quant Time: May 22 09:20:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

