

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052120\
 Data File : VX016382.D
 Acq On : 21 May 2020 14:41
 Operator : JC/SP
 Sample : L2672-08DL 200X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP021MW347-200514DL

Quant Time: May 22 09:23:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	241164	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	393921	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	382986	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	191471	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	137254	43.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.50%	
35) Dibromofluoromethane	5.47	113	112281	44.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.98%	
50) Toluene-d8	8.70	98	479521	49.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.74%	
62) 4-Bromofluorobenzene	11.12	95	192938	52.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.50%	
Target Compounds						
4) Vinyl Chloride	1.39	62	12604	3.985	ug/l	98
27) cis-1,2-Dichloroethene	4.57	96	94976	31.097	ug/l	85
44) Trichloroethene	7.19	130	12458	3.091	ug/l	95
65) Chlorobenzene	10.12	112	2314	0.289	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052120\
Data File : VX016382.D
Acq On : 21 May 2020 14:41
Operator : JC/SP
Sample : L2672-08DL 200X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP021MW347-200514DL

Quant Time: May 22 09:23:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

