

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052120\
 Data File : VX016392.D
 Acq On : 21 May 2020 18:27
 Operator : JC/SP
 Sample : L2730-05 40X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z2-DUP-0558-200520

Quant Time: May 22 09:44:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	234441	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	383414	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	378527	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	195695	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	133890	43.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.80%	
35) Dibromofluoromethane	5.47	113	109988	45.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.56%	
50) Toluene-d8	8.70	98	477675	51.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.08%	
62) 4-Bromofluorobenzene	11.12	95	192623	54.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.22%	
Target Compounds						
40) Benzene	6.11	78	7004	0.634	ug/l	98
65) Chlorobenzene	10.12	112	106973	13.497	ug/l	99
91) 1,2-Dichlorobenzene	12.38	146	4473	0.704	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052120\
Data File : VX016392.D
Acq On : 21 May 2020 18:27
Operator : JC/SP
Sample : L2730-05 40X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z2-DUP-0558-200520

Quant Time: May 22 09:44:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

