

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052120\
 Data File : VX016395.D
 Acq On : 21 May 2020 19:35
 Operator : JC/SP
 Sample : L2732-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS038-200520

Quant Time: May 22 09:48:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	226534	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	375676	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	366036	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	189498	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	131986	44.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.56%	
35) Dibromofluoromethane	5.47	113	107167	45.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.06%	
50) Toluene-d8	8.70	98	463078	50.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.00%	
62) 4-Bromofluorobenzene	11.12	95	181646	52.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.16%	
Target Compounds						
16) Acetone	2.42	43	4209	3.164	ug/l	99
27) cis-1,2-Dichloroethene	4.57	96	28323	9.873	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052120\
Data File : VX016395.D
Acq On : 21 May 2020 19:35
Operator : JC/SP
Sample : L2732-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS038-200520

Quant Time: May 22 09:48:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

