

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052121\
 Data File : VX022105.D
 Acq On : 21 May 2021 11:18
 Operator : JC/MD
 Sample : M2262-09 .2PPB MDL
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MDL-WATER-03-QT2-2021


Quant Time: May 22 06:23:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 22 06:22:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	111942	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	213765	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186185	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77125	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	92132	53.35	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 106.70%			
35) Dibromofluoromethane	5.391	113	70578	52.06	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.12%			
50) Toluene-d8	8.653	98	269411	51.59	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 103.18%			
62) 4-Bromofluorobenzene	11.085	95	94126	48.77	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.54%			
Target Compounds					Qvalue	
3) Chloromethane	1.294	50	405	0.28	ug/l #	82
4) Vinyl Chloride	1.374	62	349	0.22	ug/l #	43
5) Bromomethane	1.611	94	313	0.35	ug/l #	42
6) Chloroethane	1.697	64	240	0.26	ug/l #	43
9) 1,1,2-Trichlorotrifluo...	2.325	101	334	0.25	ug/l #	51
10) Methyl Iodide	2.459	142	401	0.28	ug/l #	84
11) Tert butyl alcohol	2.953	59	777	1.97	ug/l #	91
12) 1,1-Dichloroethene	2.325	96	260	0.20	ug/l	83
13) Acrolein	2.233	56	287	1.10	ug/l #	59
14) Allyl chloride	2.654	41	564	0.24	ug/l #	34
15) Acrylonitrile	3.069	53	568	0.73	ug/l #	78
16) Acetone	2.386	43	1012	1.35	ug/l #	78
17) Carbon Disulfide	2.508	76	841	0.25	ug/l #	75
18) Methyl Acetate	2.709	43	575	0.29	ug/l #	47
20) Methylene Chloride	2.788	84	628	0.38	ug/l #	86
23) Vinyl Acetate	3.733	43	3374	0.77	ug/l	97
25) 2-Butanone	4.568	43	1093	0.90	ug/l #	83
29) Tetrahydrofuran	5.013	42	385	0.53	ug/l #	50
38) Carbon Tetrachloride	5.562	117	12447	6.44	ug/l #	14
39) Methylcyclohexane	7.367	83	463	0.21	ug/l #	19
42) 1,2-Dichloroethane	6.086	62	465	0.22	ug/l #	73
43) Isopropyl Acetate	6.348	43	807	0.20	ug/l #	57
48) Methyl methacrylate	7.696	41	416	0.23	ug/l #	81
49) 1,4-Dioxane	7.696	88	19	0.53	ug/l #	1
51) 4-Methyl-2-Pentanone	8.580	43	1913	0.81	ug/l #	84
58) 2-Chloroethyl Vinyl ether	8.244	63	1001	0.73	ug/l	91
59) 2-Hexanone	9.433	43	1373	0.77	ug/l	75
68) m/p-Xylenes	10.311	106	959	0.36	ug/l	84
76) 1,2,3-Trichloropropane	11.085	75	48123	25.29	ug/l #	38
81) trans-1,4-Dichloro-2-b...	11.085	75	48123	59.93	ug/l #	12

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 22 06:23:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Sat May 22 06:22:48 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18




Instrument :
MSVOA_X
ClientSampleId :
MDL-WATER-03-QT2-2021


Tgt Ion:168 Resp: 111942
Ion Ratio Lower Upper
168 100
99 68.1 45.8 68.6



#3
Chloromethane
Concen: 0.28 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.001 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18

Tgt Ion: 50 Resp: 405
Ion Ratio Lower Upper
50 100
52 22.6 26.4 39.6#









#4
 Vinyl Chloride
 Concen: 0.22 ug/l
 RT: 1.374 min Scan# 47
 Delta R.T. -0.000 min
 Lab File: VX022105.D
 Acq: 21 May 2021 11:18




Instrument :
 MSVOA_X
 ClientSampleId :
 MDL-WATER-03-QT2-2021


Tgt Ion: 62 Resp: 349
 Ion Ratio Lower Upper
 62 100
 64 0.0 25.7 38.5#



#5
 Bromomethane
 Concen: 0.35 ug/l
 RT: 1.611 min Scan# 86
 Delta R.T. -0.001 min
 Lab File: VX022105.D
 Acq: 21 May 2021 11:18

Tgt Ion: 94 Resp: 313
 Ion Ratio Lower Upper
 94 100
 96 151.6 76.4 114.6#









#6
Chloroethane
Concen: 0.26 ug/l
RT: 1.697 min Scan# 100
Delta R.T. 0.012 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18




Instrument :
MSVOA_X
ClientSampleId :
MDL-WATER-03-QT2-2021


Tgt Ion: 64 Resp: 240
Ion Ratio Lower Upper
64 100
66 0.0 25.4 38.0#



#9
1,1,2-Trichlorotrifluoroethane
Concen: 0.25 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.006 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18


Tgt Ion: 101 Resp: 334
Ion Ratio Lower Upper
101 100
85 22.5 34.9 52.3#
151 26.0 61.4 92.2#








#10
Methyl Iodide
Concen: 0.28 ug/l
RT: 2.459 min Scan# 225
Delta R.T. 0.006 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18

Instrument :
MSVOA_X
ClientSampleId :
MDL-WATER-03-QT2-2021







Tgt Ion:142 Resp: 401
Ion Ratio Lower Upper
142 100
127 47.9 33.3 49.9
141 0.0 11.4 17.0#



#11
Tert butyl alcohol
Concen: 1.97 ug/l
RT: 2.953 min Scan# 306
Delta R.T. -0.036 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18

Tgt Ion: 59 Resp: 777
Ion Ratio Lower Upper
59 100
57 6.3 7.8 11.8#





#12

1,1-Dichloroethene

Concen: 0.20 ug/l

RT: 2.325 min Scan# 203

Delta R.T. 0.006 min





Lab File: VX022105.D

Acq: 21 May 2021 11:18

Tgt Ion: 96 Resp: 260

Ion	Ratio	Lower	Upper
96	100		
61	127.6	124.6	187.0
98	67.5	50.6	76.0

Instrument :
MSVOA_X
ClientSampleId :
MDL-WATER-03-QT2-2021



#13

Acrolein

Concen: 1.10 ug/l

RT: 2.233 min Scan# 188




Delta R.T. -0.007 min


Lab File: VX022105.D

Acq: 21 May 2021 11:18

Tgt Ion: 56 Resp: 287


Ion	Ratio	Lower	Upper
56	100		
55	104.5	56.4	84.6#








#14
 Allyl chloride
 Concen: 0.24 ug/l
 RT: 2.654 min Scan# 257
 Delta R.T. -0.012 min
 Lab File: VX022105.D
 Acq: 21 May 2021 11:18


Instrument :
 MSVOA_X
 ClientSampleId :
 MDL-WATER-03-QT2-2021





Tgt Ion: 41	Resp: 564
Ion Ratio	Lower Upper
41 100	
39 0.0	51.6 77.4#
76 15.2	30.5 45.7#




#15
 Acrylonitrile
 Concen: 0.73 ug/l
 RT: 3.069 min Scan# 325
 Delta R.T. -0.000 min
 Lab File: VX022105.D
 Acq: 21 May 2021 11:18



Tgt Ion: 53	Resp: 568
Ion Ratio	Lower Upper
53 100	
52 59.0	65.3 97.9#
51 26.8	28.2 42.2#









#16
Acetone
Concen: 1.35 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18




Instrument :
MSVOA_X
ClientSampleId :
MDL-WATER-03-QT2-2021


Tgt Ion: 43 Resp: 1012
Ion Ratio Lower Upper
43 100
58 22.1 27.8 41.8#



#17
Carbon Disulfide
Concen: 0.25 ug/l
RT: 2.508 min Scan# 233
Delta R.T. -0.006 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18

Tgt Ion: 76 Resp: 841
Ion Ratio Lower Upper
76 100
78 0.0 7.2 10.8#





#18

Methyl Acetate

Concen: 0.29 ug/l

RT: 2.709 min Scan# 266

Delta R.T. -0.000 min

Lab File: VX022105.D





Acq: 21 May 2021 11:18

Tgt Ion: 43 Resp: 575

Ion Ratio Lower Upper

43 100

74 0.0 21.9 32.9#



#20

Methylene Chloride

Concen: 0.38 ug/l

RT: 2.788 min Scan# 279

Delta R.T. -0.006 min

Lab File: VX022105.D

Acq: 21 May 2021 11:18

Tgt Ion: 84 Resp: 628




Ion Ratio Lower Upper


84 100

49 101.8 88.6 132.8

51 30.3 27.0 40.6

86 85.5 50.3 75.5#









#23
 Vinyl Acetate
 Concen: 0.77 ug/l
 RT: 3.733 min Scan# 434
 Delta R.T. 0.006 min
 Lab File: VX022105.D
 Acq: 21 May 2021 11:18




Instrument :
 MSVOA_X
 ClientSampleId :
 MDL-WATER-03-QT2-2021


Tgt Ion: 43 Resp: 3374
 Ion Ratio Lower Upper
 43 100
 86 11.4 9.9 14.9



#25
 2-Butanone
 Concen: 0.90 ug/l
 RT: 4.568 min Scan# 571
 Delta R.T. 0.000 min
 Lab File: VX022105.D
 Acq: 21 May 2021 11:18

Tgt Ion: 43 Resp: 1093
 Ion Ratio Lower Upper
 43 100
 72 20.2 23.7 35.5#





#29

Tetrahydrofuran

Concen: 0.53 ug/l

RT: 5.013 min Scan# 644





Delta R.T. 0.000 min

Lab File: VX022105.D

Acq: 21 May 2021 11:18

Tgt Ion:	42	Resp:	385
Ion	Ratio	Lower	Upper
42	100		
72	0.0	41.7	62.5#
71	63.6	38.4	57.6#

Instrument :
MSVOA_X
ClientSampleId :
MDL-WATER-03-QT2-2021



#33

1,2-Dichloroethane-d4

Concen: 53.35 ug/l




RT: 5.964 min Scan# 800


Delta R.T. 0.000 min

Lab File: VX022105.D

Acq: 21 May 2021 11:18

Tgt Ion:	65	Resp:	92132
Ion	Ratio	Lower	Upper
65	100		
67	52.0	0.0	105.4









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18




Instrument :
MSVOA_X
ClientSampleId :
MDL-WATER-03-QT2-2021


Tgt Ion:114 Resp: 213765
Ion Ratio Lower Upper
114 100
63 20.9 0.0 40.8
88 16.0 0.0 32.8



#35
Dibromofluoromethane
Concen: 52.06 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18


Tgt Ion:113 Resp: 70578
Ion Ratio Lower Upper
113 100
111 102.6 81.9 122.9
192 16.0 16.7 25.1#








#38
Carbon Tetrachloride
Concen: 6.44 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.128 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18

Instrument :
MSVOA_X
ClientSampleId :
MDL-WATER-03-QT2-2021







Tgt Ion:117 Resp: 12447
Ion Ratio Lower Upper
117 100
119 2.7 76.2 114.2#
121 0.0 24.8 37.2#



#39
Methylcyclohexane
Concen: 0.21 ug/l
RT: 7.367 min Scan# 1030
Delta R.T. -0.018 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18

Tgt Ion: 83 Resp: 463
Ion Ratio Lower Upper
83 100
55 0.0 58.9 88.3#
98 0.0 38.1 57.1#





#42

1,2-Dichloroethane

Concen: 0.22 ug/l

RT: 6.086 min Scan# 820

Delta R.T. -0.013 min

Lab File: VX022105.D





Acq: 21 May 2021 11:18

Tgt Ion: 62 Resp: 465

Ion Ratio Lower Upper

62 100

98 0.0 0.0 20.2



#43

Isopropyl Acetate

Concen: 0.20 ug/l

RT: 6.348 min Scan# 863

Delta R.T. -0.001 min

Lab File: VX022105.D

Acq: 21 May 2021 11:18




Tgt Ion: 43 Resp: 807


Ion Ratio Lower Upper

43 100

61 0.0 18.2 27.4#

87 0.0 12.5 18.7#









#48
Methyl methacrylate
Concen: 0.23 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18




Instrument :
MSVOA_X
ClientSampleId :
MDL-WATER-03-QT2-2021


Tgt Ion: 41 Resp: 416
Ion Ratio Lower Upper
41 100
69 66.8 76.0 114.0#
39 56.3 44.6 67.0



#49
1,4-Dioxane
Concen: 0.53 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.012 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18

Tgt Ion: 88 Resp: 19
Ion Ratio Lower Upper
88 100
43 0.0 23.2 34.8#
58 752.6 55.0 82.4#









#50
Toluene-d8
Concen: 51.59 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18




Instrument :
MSVOA_X
ClientSampleId :
MDL-WATER-03-QT2-2021


Tgt Ion: 98 Resp: 269411
Ion Ratio Lower Upper
98 100
100 66.4 51.9 77.9



#51
4-Methyl-2-Pentanone
Concen: 0.81 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. -0.000 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18

Tgt Ion: 43 Resp: 1913
Ion Ratio Lower Upper
43 100
58 34.1 35.4 53.0#









#58
2-Chloroethyl Vinyl ether
Concen: 0.73 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. -0.007 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18




Instrument :
MSVOA_X
ClientSampleId :
MDL-WATER-03-QT2-2021


Tgt Ion: 63 Resp: 1001
Ion Ratio Lower Upper
63 100
106 34.4 23.5 35.3



#59
2-Hexanone
Concen: 0.77 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18

Tgt Ion: 43 Resp: 1373
Ion Ratio Lower Upper
43 100
58 43.0 30.9 92.8









#62
4-Bromofluorobenzene
Concen: 48.77 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18




Instrument :
MSVOA_X
ClientSampleId :
MDL-WATER-03-QT2-2021


Tgt Ion: 95 Resp: 94126
Ion Ratio Lower Upper
95 100
174 66.4 0.0 154.6
176 63.2 0.0 155.8



#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18

Tgt Ion: 117 Resp: 186185
Ion Ratio Lower Upper
117 100
82 60.7 47.4 71.2
119 32.3 26.6 39.8









#68
m/p-Xylenes
Concen: 0.36 ug/l
RT: 10.311 min Scan# 1513
Delta R.T. 0.006 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18




Instrument :
MSVOA_X
ClientSampleId :
MDL-WATER-03-QT2-2021


Tgt Ion:106 Resp: 959
Ion Ratio Lower Upper
106 100
91 178.9 162.8 244.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18

Tgt Ion:152 Resp: 77125
Ion Ratio Lower Upper
152 100
115 66.9 41.1 123.3
150 158.7 0.0 349.0









#76
1,2,3-Trichloropropane
Concen: 25.29 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. -0.159 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18

Instrument :
MSVOA_X
ClientSampleId :
MDL-WATER-03-QT2-2021

Tgt Ion: 75 Resp: 48123
Ion Ratio Lower Upper
75 100
77 1.0 19.1 57.5#



#81
trans-1,4-Dichloro-2-butene
Concen: 59.93 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.060 min
Lab File: VX022105.D
Acq: 21 May 2021 11:18

Tgt Ion: 75 Resp: 48123
Ion Ratio Lower Upper
75 100
53 0.0 74.5 111.7#
89 0.0 37.9 56.9#

