

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052121\
 Data File : VX022114.D
 Acq On : 21 May 2021 14:50
 Operator : JC/MD
 Sample : M2472-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW8-GW-898-900

Quant Time: May 22 04:54:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	74462	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	148025	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	130643	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51405	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	57928	50.43	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.86%
35) Dibromofluoromethane	5.397	113	46619	49.66	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.32%
50) Toluene-d8	8.653	98	179357	49.60	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.20%
62) 4-Bromofluorobenzene	11.085	95	62474	46.74	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.48%
Target Compounds						
16) Acetone	2.386	43	16337	32.66	ug/l	94
25) 2-Butanone	4.574	43	4190	5.18	ug/l #	83
40) Benzene	6.037	78	2635	0.63	ug/l	93
65) Chlorobenzene	10.079	112	1065	0.39	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052121\
Data File : VX022114.D
Acq On : 21 May 2021 14:50
Operator : JC/MD
Sample : M2472-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-898-900

Quant Time: May 22 04:54:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

