

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052121\
 Data File : VX022115.D
 Acq On : 21 May 2021 15:11
 Operator : JC/MD
 Sample : M2472-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW8-GW-918-920

Quant Time: May 22 04:55:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	70283	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	142348	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	119639	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43716	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	56056	51.70	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.40%
35) Dibromofluoromethane	5.404	113	44344	49.12	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.24%
50) Toluene-d8	8.653	98	170889	49.14	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.28%
62) 4-Bromofluorobenzene	11.085	95	57442	44.69	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.38%
Target Compounds						
16) Acetone	2.386	43	2941	6.23	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052121\
Data File : VX022115.D
Acq On : 21 May 2021 15:11
Operator : JC/MD
Sample : M2472-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-918-920

Quant Time: May 22 04:55:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

