

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052121\
 Data File : VX022122.D
 Acq On : 21 May 2021 17:50
 Operator : JC/MD
 Sample : M2406-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: May 22 04:57:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	69203	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	140586	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	121262	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47022	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	57831	54.17	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.34%
35) Dibromofluoromethane	5.397	113	45648	51.20	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.40%
50) Toluene-d8	8.653	98	177538	51.70	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.40%
62) 4-Bromofluorobenzene	11.085	95	61344	48.33	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.66%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052121\
Data File : VX022122.D
Acq On : 21 May 2021 17:50
Operator : JC/MD
Sample : M2406-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: May 22 04:57:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

