

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052121\
 Data File : VX022125.D
 Acq On : 21 May 2021 19:01
 Operator : JC/MD
 Sample : M2406-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: May 22 04:58:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	66830	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	137159	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	120573	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47022	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	57107	55.40	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.80%
35) Dibromofluoromethane	5.397	113	44204	50.82	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.64%
50) Toluene-d8	8.653	98	173692	51.84	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.68%
62) 4-Bromofluorobenzene	11.085	95	60438	48.80	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.60%

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052121\
Data File : VX022125.D
Acq On : 21 May 2021 19:01
Operator : JC/MD
Sample : M2406-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: May 22 04:58:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

