

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052121\
 Data File : VX022126.D
 Acq On : 21 May 2021 19:24
 Operator : JC/MD
 Sample : PB136366TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136366TB

Quant Time: May 22 04:58:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	64242	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	136036	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	118820	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46023	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	52672	53.15	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.30%
35) Dibromofluoromethane	5.397	113	41555	48.17	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.34%
50) Toluene-d8	8.653	98	165583	49.83	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.66%
62) 4-Bromofluorobenzene	11.085	95	57486	46.80	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.60%
Target Compounds						
16) Acetone	2.392	43	15659	36.29	ug/l	95
20) Methylene Chloride	2.794	84	7111	7.42	ug/l	93
43) Isopropyl Acetate	6.355	43	13587	5.39	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052121\
Data File : VX022126.D
Acq On : 21 May 2021 19:24
Operator : JC/MD
Sample : PB136366TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136366TB

Quant Time: May 22 04:58:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

