

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052121\
 Data File : VX022127.D
 Acq On : 21 May 2021 19:48
 Operator : JC/MD
 Sample : PB136552TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136552TB

Quant Time: May 22 04:58:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	65169	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	134949	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	118864	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45902	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52697	52.42	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.84%
35) Dibromofluoromethane	5.397	113	41360	48.33	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.66%
50) Toluene-d8	8.653	98	166059	50.37	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.74%
62) 4-Bromofluorobenzene	11.085	95	55886	45.87	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.74%
Target Compounds						
16) Acetone	2.392	43	12121	27.69	ug/l	Qvalue 100
18) Methyl Acetate	2.709	43	1618	1.39	ug/l	# 86
20) Methylene Chloride	2.794	84	6988	7.19	ug/l	96
43) Isopropyl Acetate	6.355	43	14370	5.74	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052121\
Data File : VX022127.D
Acq On : 21 May 2021 19:48
Operator : JC/MD
Sample : PB136552TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136552TB

Quant Time: May 22 04:58:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

