

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052121\
 Data File : VX022128.D
 Acq On : 21 May 2021 20:12
 Operator : JC/MD
 Sample : M2467-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW-BR3-SO-051921

Quant Time: May 22 04:58:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	62544	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	130505	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	110526	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39803	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	50998	52.86	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.72%
35) Dibromofluoromethane	5.403	113	40314	48.71	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.42%
50) Toluene-d8	8.653	98	156085	48.96	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.92%
62) 4-Bromofluorobenzene	11.085	95	51371	43.60	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.20%
Target Compounds						
16) Acetone	2.385	43	15277	36.37	ug/l	97
20) Methylene Chloride	2.794	84	7255	7.78	ug/l #	81
43) Isopropyl Acetate	6.348	43	12543	5.18	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052121\
Data File : VX022128.D
Acq On : 21 May 2021 20:12
Operator : JC/MD
Sample : M2467-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW-BR3-SO-051921

Quant Time: May 22 04:58:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

