

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052124\
 Data File : VX041580.D
 Acq On : 21 May 2024 11:23
 Operator : JC/MD
 Sample : VX0521WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0521WBL01

Quant Time: May 22 02:52:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 16 05:13:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	28005	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.757	114	155217	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	139140	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	65528	29.713	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.033%
60) 4-Bromofluorobenzene	11.079	95	59677	27.475	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.567%
63) Toluene-d8	8.647	98	177410	29.745	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.133%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052124\
Data File : VX041580.D
Acq On : 21 May 2024 11:23
Operator : JC/MD
Sample : VX0521WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0521WBL01

Quant Time: May 22 02:52:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 16 05:13:13 2024
Response via : Initial Calibration

