

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052124\
 Data File : VX041582.D
 Acq On : 21 May 2024 12:29
 Operator : JC/MD
 Sample : P2546-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 240516002-05-TRIP-BLANK

Quant Time: May 22 02:53:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 16 05:13:13 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 05/22/2024
 Supervised By :Semsettin Yesilyurt 05/22/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	27055	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	147657	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	131285	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	64634	30.337	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.133%
60) 4-Bromofluorobenzene	11.079	95	56698	27.665	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	92.200%
63) Toluene-d8	8.647	98	169429	30.106	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.367%
Target Compounds						
15) Acetone	2.386	58	1472	5.994	ug/l	Qvalue # 68
25) Chloroform	5.105	83	5484m	1.777	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052124\
Data File : VX041582.D
Acq On : 21 May 2024 12:29
Operator : JC/MD
Sample : P2546-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
240516002-05-TRIP-BLANK

Quant Time: May 22 02:53:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 16 05:13:13 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 05/22/2024
Supervised By :Semsettin Yesilyurt 05/22/2024

