



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052124\
Data File : VX041576.D
Acq On : 21 May 2024 08:39
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Thu May 16 05:13:13 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.8	9707	PASS
75	95	30	60	51.2	25099	PASS
95	95	100	100	100.0	49043	PASS
96	95	5	9	6.5	3212	PASS
173	174	0.00	2	0.3	164	PASS
174	95	50	100	95.7	46915	PASS
175	174	5	9	7.8	3649	PASS
176	174	95	101	97.4	45712	PASS
177	176	5	9	6.5	2958	PASS