

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052125\
 Data File : VX046285.D
 Acq On : 21 May 2025 10:22
 Operator : JC/MD
 Sample : VX0521MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0521MBL01

Quant Time: May 22 01:45:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	63482	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	122597	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	113349	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	46128	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61333	51.823	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.640%
35) Dibromofluoromethane	5.379	113	45339	51.357	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.720%
50) Toluene-d8	8.646	98	152832	50.017	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.040%
62) 4-Bromofluorobenzene	11.079	95	56671	48.351	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052125\
Data File : VX046285.D
Acq On : 21 May 2025 10:22
Operator : JC/MD
Sample : VX0521MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0521MBL01

Quant Time: May 22 01:45:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

