

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052125\
 Data File : VX046296.D
 Acq On : 21 May 2025 14:42
 Operator : JC/MD
 Sample : Q2082-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VPB182-HYD-20250516

Quant Time: May 22 01:50:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	67553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	134394	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	128060	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54802	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65952	52.368	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.740%	
35) Dibromofluoromethane	5.385	113	48556	50.173	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.340%	
50) Toluene-d8	8.647	98	165672	49.460	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.920%	
62) 4-Bromofluorobenzene	11.079	95	66559	51.802	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.600%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1819	3.602	ug/l	93
47) Bromodichloromethane	7.830	83	915	0.622	ug/l #	91
60) Dibromochloromethane	9.525	129	1211	1.197	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052125\
Data File : VX046296.D
Acq On : 21 May 2025 14:42
Operator : JC/MD
Sample : Q2082-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VPB182-HYD-20250516

Quant Time: May 22 01:50:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

