

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052125\
 Data File : VX046301.D
 Acq On : 21 May 2025 16:39
 Operator : JC/MD
 Sample : Q2082-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-390-392

Quant Time: May 22 01:51:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	65898	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	130017	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	123419	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52601	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66464	54.100	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.200%			
35) Dibromofluoromethane	5.385	113	47621	50.863	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.720%			
50) Toluene-d8	8.647	98	164758	50.843	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.680%			
62) 4-Bromofluorobenzene	11.079	95	64090	51.560	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 103.120%			
Target Compounds						
16) Acetone	2.380	43	1249	2.536	ug/l #	86
17) Carbon Disulfide	2.508	76	910	0.491	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052125\
Data File : VX046301.D
Acq On : 21 May 2025 16:39
Operator : JC/MD
Sample : Q2082-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-390-392

Quant Time: May 22 01:51:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

