

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052219\
 Data File : VX009840.D
 Acq On : 23 May 2019 06:21
 Operator : JC/SP
 Sample : K2976-21 50X
 Misc : 6.62g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EP1-SBR-4BR(14-17)

Quant Time: May 23 07:26:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	183305	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	298490	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	282777	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	145634	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	125932	50.33	ug/l	0.00
Spiked Amount			Recovery	=	100.66%	
35) Dibromofluoromethane	5.50	113	90419	48.19	ug/l	0.00
Spiked Amount			Recovery	=	96.38%	
50) Toluene-d8	8.71	98	384762	50.79	ug/l	0.00
Spiked Amount			Recovery	=	101.58%	
62) 4-Bromofluorobenzene	11.13	95	149778	54.41	ug/l	0.00
Spiked Amount			Recovery	=	108.82%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.04	59	125	0.216	ug/l #	26
16) Acetone	2.45	43	473	0.352	ug/l #	62
28) Bromochloromethane	4.99	49	24	Below Cal	#	20
31) Cyclohexane	5.66	56	4390	1.187	ug/l #	1
36) 1,1-Dichloropropene	5.66	75	14418	5.198	ug/l #	50
38) Carbon Tetrachloride	5.66	117	17724	7.033	ug/l #	15
55) 1,1,2-Trichloroethane	9.16	97	5241	2.462	ug/l #	20
56) Ethyl methacrylate	9.16	69	1447	0.381	ug/l #	1
59) 2-Hexanone	9.45	43	36353	11.455	ug/l #	31
73) Isopropylbenzene	11.02	105	2910	0.275	ug/l	100
74) N-amyl acetate	10.91	43	79828	12.929	ug/l #	23
75) 1,1,2,2-Tetrachloroethane	11.27	83	57837	14.650	ug/l #	25
76) 1,2,3-Trichloropropane	11.13	75	76188	22.553	ug/l #	36
78) n-propylbenzene	11.36	91	14135	1.169	ug/l	92
79) 2-Chlorotoluene	11.44	91	2745	0.368	ug/l	77
81) trans-1,4-Dichloro-2-buten	11.13	75	76188	57.705	ug/l #	16
83) tert-Butylbenzene	11.77	119	19455	2.243	ug/l	82
84) 1,2,4-Trimethylbenzene	11.80	105	95058	10.620	ug/l	100
85) sec-Butylbenzene	11.94	105	129498	12.592	ug/l	98
86) p-Isopropyltoluene	12.02	119	214069	23.154	ug/l	97
89) n-Butylbenzene	12.37	91	25807	3.088	ug/l #	1
90) Hexachloroethane	12.58	117	21145	13.424	ug/l #	7
92) 1,2-Dibromo-3-Chloropropan	12.97	75	1588	1.793	ug/l #	2


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052219\
Data File : VX009840.D
Acq On : 23 May 2019 06:21
Operator : JC/SP
Sample : K2976-21 50X
Misc : 6.62g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EP1-SBR-4BR(14-17)

Quant Time: May 23 07:26:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21

Instrument :

MSVOA_X

ClientSampleId :


EP1-SBR-4BR(14-17)

Tot Ion:168 Resp: 183305

Ion Ratio Lower Upper




168 100

99 57.8 42.3 63.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#11

Tert butyl alcohol

Concen: 0.216 ug/l

RT: 3.04 min Scan# 309

Delta R.T. -0.01 min

Lab File: VX009840.D


Acq: 23 May 2019 06:21

Tgt Ion: 59 Resp: 125

Ion Ratio Lower Upper



59 100


57 37.6 8.2 12.2#



Abundance Ion 59.00 (58.70 to 59.70): VX0

Ion 57.00 (56.70 to 57.70): VX0





#16

Acetone

Concen: 0.352 ug/l

RT: 2.45 min Scan# 213

Delta R.T. 0.01 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21

Instrument :

MSVOA_X

ClientSampleId :





EP1-SBR-4BR(14-17)

Tot Ion: 43 Resp: 473

Ion Ratio Lower Upper

43 100

58 52.2 24.9 37.3#



#28

Bromochloromethane

Concen: Below Cal

RT: 4.99 min Scan# 630

Delta R.T. -0.02 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21




Tgt Ion: 49 Resp: 24


Ion Ratio Lower Upper

49 100

129 0.0 0.0 3.8

130 0.0 51.2 76.8#









#31
Cyclohexane
Concen: 1.187 ug/l
RT: 5.66 min Scan# 739
Delta R.T. 0.08 min
Lab File: VX009840.D
Acq: 23 May 2019 06:21




Instrument :
MSVOA_X
ClientSampleId :
EP1-SBR-4BR(14-17)


Tot Ion: 56 Resp: 4390
Ion Ratio Lower Upper
56 100
69 171.4 23.4 35.0#
84 141.7 61.9 92.9#



#33
1,2-Dichloroethane-d4
Concen: 50.326 ug/l
RT: 6.06 min Scan# 805
Delta R.T. 0.00 min
Lab File: VX009840.D
Acq: 23 May 2019 06:21

Tgt Ion: 65 Resp: 125932
Ion Ratio Lower Upper
65 100
67 53.0 0.0 106.6









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. 0.00 min
 Lab File: VX009840.D
 Acq: 23 May 2019 06:21

Instrument :
 MSVOA_X
 ClientSampleId :
 EP1-SBR-4BR(14-17)

Tot Ion:114 Resp: 298490
 Ion Ratio Lower Upper
 114 100
 63 23.3 0.0 45.6
 88 17.1 0.0 33.2




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX0
 Ion 87.90 (87.60 to 88.60): VX0






#35
 Dibromofluoromethane
 Concen: 48.187 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. 0.01 min
 Lab File: VX009840.D
 Acq: 23 May 2019 06:21

Tgt Ion:113 Resp: 90419
 Ion Ratio Lower Upper
 113 100
 111 103.2 82.8 124.2
 192 21.6 17.0 25.6



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 5.198 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.14 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21

Instrument :

MSVOA_X

ClientSampleId :

EP1-SBR-4BR(14-17)


Tot Ion: 75 Resp: 14418

Ion Ratio Lower Upper

75 100

110 8.8 16.9 50.6#




77 0.0 24.8 37.2#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VX0



#38

Carbon Tetrachloride

Concen: 7.033 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.12 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21


Tgt Ion: 117 Resp: 17724

Ion Ratio Lower Upper

117 100

119 4.7 77.5 116.3#



121 0.0 24.7 37.1#




Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V





#50

Toluene-d8

Concen: 50.790 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21

Instrument :

MSVOA_X

ClientSampleId :


EP1-SBR-4BR(14-17)

Tot Ion: 98 Resp: 384762

Ion Ratio Lower Upper




98 100

100 64.3 51.2 76.8



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX0



#55

1,1,2-Trichloroethane

Concen: 2.462 ug/l

RT: 9.16 min Scan# 1314

Delta R.T. -0.05 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21

Tgt Ion: 97 Resp: 5241


Ion Ratio Lower Upper

97 100

83 9.8 68.9 103.3#

85 0.0 43.8 65.6#

99 0.0 48.2 72.2#






Abundance Ion 96.90 (96.60 to 97.60): VX0

Ion 82.90 (82.60 to 83.60): VX0

Ion 84.90 (84.60 to 85.60): VX0

Ion 98.90 (98.60 to 99.60): VX0





#56

Ethyl methacrylate

Concen: 0.381 ug/l

RT: 9.16 min Scan# 1314

Delta R.T. -0.01 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21

Instrument :

MSVOA_X

ClientSampleId :

EP1-SBR-4BR(14-17)


Tot Ion: 69 Resp: 1447

Ion Ratio Lower Upper

69 100

41 156.3 60.6 90.8#




39 118.0 28.2 42.4#



Abundance Ion 69.00 (68.70 to 69.70): VX0

Ion 41.00 (40.70 to 41.70): VX0

Ion 39.00 (38.70 to 39.70): VX0



#59

2-Hexanone

Concen: 11.455 ug/l

RT: 9.45 min Scan# 1360

Delta R.T. -0.04 min

Lab File: VX009840.D


Acq: 23 May 2019 06:21

Tgt Ion: 43 Resp: 36353

Ion Ratio Lower Upper



43 100


58 3.5 25.9 77.7#



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0





#62

4-Bromofluorobenzene

Concen: 54.412 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21

Instrument :

MSVOA_X

ClientSampleId :

EP1-SBR-4BR(14-17)


Tot Ion: 95 Resp: 149778

Ion Ratio Lower Upper

95 100

174 78.8 0.0 161.6




176 76.5 0.0 159.2



Abundance Ion 94.90 (94.60 to 95.60): VX009840.D (-1588) (-)

Ion 173.80 (173.50 to 174.50): VX009840.D (-1588) (-)

Ion 175.80 (175.50 to 176.50): VX009840.D (-1588) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.11 min Scan# 1469

Delta R.T. 0.00 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21


Tgt Ion: 117 Resp: 282777

Ion Ratio Lower Upper

117 100

82 61.0 48.8 73.2



119 32.2 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): VX009840.D (-1420) (-)

Ion 82.00 (81.70 to 82.70): VX009840.D (-1420) (-)

Ion 118.90 (118.60 to 119.60): VX009840.D (-1420) (-)





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. 0.00 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21

Instrument :

MSVOA_X

ClientSampleId :

EP1-SBR-4BR(14-17)


Tot Ion:152 Resp: 145634

Ion Ratio Lower Upper

152 100

115 58.4 41.2 123.6




150 156.4 0.0 346.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#73

Isopropylbenzene

Concen: 0.275 ug/l

RT: 11.02 min Scan# 1618

Delta R.T. 0.00 min

Lab File: VX009840.D


Acq: 23 May 2019 06:21

Tgt Ion:105 Resp: 2910

Ion Ratio Lower Upper



105 100


120 25.4 12.8 38.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#74

N-amvl acetate

Concen: 12.929 ug/l

RT: 10.91 min Scan# 1600

Delta R.T. 0.01 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21

Instrument :

MSVOA_X


ClientSampleId :

EP1-SBR-4BR(14-17)

Tot Ion: 43 Resp: 79828

Ion Ratio Lower Upper

43	100		
70	32.0	30.0	45.0
55	128.5	17.9	26.9#
61	0.0	18.4	27.6#






Abundance

Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0



#75

1,1,2,2-Tetrachloroethane

Concen: 14.650 ug/l

RT: 11.27 min Scan# 1660

Delta R.T. 0.01 min


Lab File: VX009840.D

Acq: 23 May 2019 06:21

Tgt Ion: 83 Resp: 57837

Ion Ratio Lower Upper

83	100		
131	0.0	4.6	13.8#
85	0.0	32.0	96.2#






Abundance

Ion 82.90 (82.60 to 83.60): VX0

Ion 130.80 (130.50 to 131.50): V

Ion 84.90 (84.60 to 85.60): VX0









#76
 1,2,3-Trichloropropane
 Concen: 22.553 ug/l
 RT: 11.13 min Scan# 1637
 Delta R.T. -0.16 min
 Lab File: VX009840.D
 Acq: 23 May 2019 06:21




Instrument :
 MSVOA_X
 ClientSampleId :
 EP1-SBR-4BR(14-17)


Tot Ion: 75 Resp: 76188
 Ion Ratio Lower Upper
 75 100
 77 2.6 22.1 66.1#



#78
 n-propylbenzene
 Concen: 1.169 ug/l
 RT: 11.36 min Scan# 1674
 Delta R.T. 0.00 min
 Lab File: VX009840.D
 Acq: 23 May 2019 06:21

Tgt Ion: 91 Resp: 14135
 Ion Ratio Lower Upper
 91 100
 120 18.0 11.0 33.0









#79
 2-Chlorotoluene
 Concen: 0.368 ug/l
 RT: 11.44 min Scan# 1688
 Delta R.T. 0.02 min
 Lab File: VX009840.D
 Acq: 23 May 2019 06:21




Instrument :
 MSVOA_X
 ClientSampleId :
 EP1-SBR-4BR(14-17)


Tot Ion: 91 Resp: 2745
 Ion Ratio Lower Upper
 91 100
 126 45.7 16.3 48.9



#81
 trans-1,4-Dichloro-2-butene
 Concen: 57.705 ug/l
 RT: 11.13 min Scan# 1637
 Delta R.T. 0.06 min
 Lab File: VX009840.D
 Acq: 23 May 2019 06:21

Tgt Ion: 75 Resp: 76188
 Ion Ratio Lower Upper
 75 100
 53 8.6 79.0 118.6#
 89 0.0 34.7 52.1#





#83

tert-Butylbenzene

Concen: 2.243 ug/l

RT: 11.77 min Scan# 1741

Delta R.T. 0.00 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21

Instrument :

MSVOA_X

ClientSampleId :

EP1-SBR-4BR(14-17)


Tot Ion:119 Resp: 19455

Ion Ratio Lower Upper

119 100

91 76.7 29.5 88.5




134 24.4 11.4 34.1



Abundance Ion 119.00 (118.70 to 119.70): V

Ion 91.00 (90.70 to 91.70): VX0

Ion 134.00 (133.70 to 134.70): V



#84

1,2,4-Trimethylbenzene

Concen: 10.620 ug/l

RT: 11.80 min Scan# 1747

Delta R.T. 0.00 min

Lab File: VX009840.D


Acq: 23 May 2019 06:21

Tgt Ion:105 Resp: 95058

Ion Ratio Lower Upper



105 100


120 43.7 22.0 66.0



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#85

sec-Butylbenzene

Concen: 12.592 ug/l

RT: 11.94 min Scan# 1770

Delta R.T. 0.00 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21

Instrument :

MSVOA_X

ClientSampleId :


EP1-SBR-4BR(14-17)

Tot Ion:105 Resp: 129498

Ion Ratio Lower Upper

105 100

134 20.0 9.7 28.9



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V

11.94

120000

100000

80000

60000



40000

20000

0

Time-->

11.85 11.90 11.95 12.00



#86

p-Isopropyltoluene

Concen: 23.154 ug/l

RT: 12.02 min Scan# 1783

Delta R.T. -0.04 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21


Tgt Ion:119 Resp: 214069

Ion Ratio Lower Upper

119 100

134 27.7 12.9 38.7

91 25.2 11.8 35.4



Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): VX0

250000

200000

150000


100000


50000

0

Time-->

11.95 12.00 12.05





#89

n-Butylbenzene

Concen: 3.088 ug/l

RT: 12.37 min Scan# 1839

Delta R.T. -0.02 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21

Instrument :

MSVOA_X

ClientSampleId :

EP1-SBR-4BR(14-17)





Tot Ion: 91 Resp: 25807

Ion Ratio Lower Upper

91 100

92 84.9 26.5 79.5#

134 242.0 12.4 37.2#



#90

Hexachloroethane

Concen: 13.424 ug/l

RT: 12.58 min Scan# 1875

Delta R.T. -0.01 min

Lab File: VX009840.D




Acq: 23 May 2019 06:21




Tgt Ion: 117 Resp: 21145

Ion Ratio Lower Upper

117 100

201 0.0 42.3 126.9#





#92

1,2-Dibromo-3-Chloropropane

Concen: 1.793 ug/l

RT: 12.97 min Scan# 1939

Delta R.T. -0.02 min

Lab File: VX009840.D

Acq: 23 May 2019 06:21

Instrument :

MSVOA_X

ClientSampleId :

EP1-SBR-4BR(14-17)

Tot Ion: 75 Resp: 1588

Ion Ratio Lower Upper

75 100

155 0.0 42.0 126.0#

157 0.0 53.2 159.6#

