

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052219\
 Data File : VX009796.D
 Acq On : 22 May 2019 12:03
 Operator : JC/SP
 Sample : PB119942TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB119942TB

Quant Time: May 23 07:28:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	166435	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	265403	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	240640	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	104665	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	111865	49.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.48%	
35) Dibromofluoromethane	5.49	113	81866	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
50) Toluene-d8	8.71	98	340639	50.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.14%	
62) 4-Bromofluorobenzene	11.13	95	113900	46.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.08%	
Target Compounds						
16) Acetone	2.44	43	7996	6.554	ug/l	99
43) Isopropyl Acetate	6.45	43	12136	2.208	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052219\
Data File : VX009796.D
Acq On : 22 May 2019 12:03
Operator : JC/SP
Sample : PB119942TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB119942TB

Quant Time: May 23 07:28:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

