

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052219\
 Data File : VX009826.D
 Acq On : 23 May 2019 00:54
 Operator : JC/SP
 Sample : K2976-05 50X
 Misc : 5.39g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EP1-SB-E-6R(14-17)

Quant Time: May 23 08:42:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	179013	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	292125	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	270079	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	132256	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	124094	50.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.56%	
35) Dibromofluoromethane	5.50	113	88653	48.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.56%	
50) Toluene-d8	8.71	98	375310	50.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.24%	
62) 4-Bromofluorobenzene	11.13	95	142628	52.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.88%	
Target Compounds						
73) Isopropylbenzene	11.02	105	73002	7.591	ug/l	99
78) n-propylbenzene	11.36	91	205025	18.666	ug/l	99
83) tert-Butylbenzene	11.77	119	13653	1.733	ug/l	83
85) sec-Butylbenzene	11.94	105	98007	10.494	ug/l	96
89) n-Butylbenzene	12.38	91	63364	8.350	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052219\
Data File : VX009826.D
Acq On : 23 May 2019 00:54
Operator : JC/SP
Sample : K2976-05 50X
Misc : 5.39g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EP1-SB-E-6R(14-17)

Quant Time: May 23 08:42:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

