

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052219\
 Data File : VX009829.D
 Acq On : 23 May 2019 02:04
 Operator : JC/SP
 Sample : K2976-08 50X
 Misc : 6.71g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EP1-SB-E-7R(17-20)

Quant Time: May 31 07:53:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	178238	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	293713	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	267637	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	131069	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	123103	50.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.18%	
35) Dibromofluoromethane	5.50	113	87939	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
50) Toluene-d8	8.71	98	372398	49.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.92%	
62) 4-Bromofluorobenzene	11.13	95	141804	52.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.70%	
Target Compounds						
73) Isopropylbenzene	11.02	105	78631	8.250	ug/l	99
78) n-propylbenzene	11.36	91	240686	22.112	ug/l	98
83) tert-Butylbenzene	11.77	119	17461	2.237	ug/l	87
85) sec-Butylbenzene	11.94	105	128843	13.920	ug/l	95
89) n-Butylbenzene	12.38	91	93926	12.489	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052219\
Data File : VX009829.D
Acq On : 23 May 2019 02:04
Operator : JC/SP
Sample : K2976-08 50X
Misc : 6.71g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EP1-SB-E-7R(17-20)

Quant Time: May 31 07:53:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

