


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX052219\
 Data File : VX009817.D
 Acq On : 22 May 2019 20:37
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Title : SW846 8260
 Last Update : Thu May 09 07:31:15 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	6676	PASS
75	95	30	60	52.0	16417	PASS
95	95	100	100	100.0	31600	PASS
96	95	5	9	6.8	2142	PASS
173	174	0.00	2	0.7	161	PASS
174	95	50	100	77.7	24561	PASS
175	174	5	9	8.0	1960	PASS
176	174	95	101	97.7	24004	PASS
177	176	5	9	6.8	1628	PASS