

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052220\
 Data File : VX016403.D
 Acq On : 22 May 2020 11:50
 Operator : JC/SP
 Sample : L2730-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP021MW351-200520

Quant Time: May 23 06:39:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	231511	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	382536	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	372457	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	187999	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	131165	43.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.12%	
35) Dibromofluoromethane	5.46	113	109648	45.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.48%	
50) Toluene-d8	8.70	98	464185	49.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.42%	
62) 4-Bromofluorobenzene	11.12	95	184468	51.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.88%	
Target Compounds						
16) Acetone	2.42	43	2677	1.969	ug/l	98
27) cis-1,2-Dichloroethene	4.57	96	3311	1.129	ug/l	82
44) Trichloroethene	7.19	130	2424	0.619	ug/l	95
64) Tetrachloroethene	9.32	164	4757	1.040	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052220\
Data File : VX016403.D
Acq On : 22 May 2020 11:50
Operator : JC/SP
Sample : L2730-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP021MW351-200520

Quant Time: May 23 06:39:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

