

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052220\
 Data File : VX016406.D
 Acq On : 22 May 2020 12:57
 Operator : JC/SP
 Sample : L2728-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 239998

Quant Time: May 23 02:25:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	224414	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	371476	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	369549	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	183071	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	129450	43.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.68%	
35) Dibromofluoromethane	5.47	113	108873	46.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.52%	
50) Toluene-d8	8.70	98	465058	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
62) 4-Bromofluorobenzene	11.12	95	183433	53.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.36%	
Target Compounds						
16) Acetone	2.42	43	13728	10.419	ug/l	99
20) Methylene Chloride	2.84	84	8742	3.218	ug/l	89
43) Isopropyl Acetate	6.42	43	59832	8.848	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052220\
Data File : VX016406.D
Acq On : 22 May 2020 12:57
Operator : JC/SP
Sample : L2728-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
239998

Quant Time: May 23 02:25:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

