

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052220\
 Data File : VX016407.D
 Acq On : 22 May 2020 13:20
 Operator : JC/SP
 Sample : L2728-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 240000

Quant Time: May 23 02:28:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	223728	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	367079	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	359219	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	184151	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	128055	43.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.00%	
35) Dibromofluoromethane	5.47	113	107603	46.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.54%	
50) Toluene-d8	8.70	98	455799	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
62) 4-Bromofluorobenzene	11.12	95	179548	52.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.36%	
Target Compounds						
16) Acetone	2.42	43	12378	9.423	ug/l	90
20) Methylene Chloride	2.84	84	8257	3.049	ug/l	95
43) Isopropyl Acetate	6.42	43	60356	9.032	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052220\
Data File : VX016407.D
Acq On : 22 May 2020 13:20
Operator : JC/SP
Sample : L2728-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
240000

Quant Time: May 23 02:28:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

