

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052220\
 Data File : VX016410.D
 Acq On : 22 May 2020 14:28
 Operator : JC/SP
 Sample : L2718-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 72-12017

Quant Time: May 23 02:38:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	217361	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	364614	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	362321	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	179948	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	124362	43.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.96%	
35) Dibromofluoromethane	5.47	113	106482	46.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.20%	
50) Toluene-d8	8.70	98	453637	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
62) 4-Bromofluorobenzene	11.12	95	178230	52.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.30%	
Target Compounds						
16) Acetone	2.42	43	18564	14.546	ug/l	94
20) Methylene Chloride	2.84	84	9491	3.608	ug/l	93
43) Isopropyl Acetate	6.42	43	57924	8.727	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052220\
Data File : VX016410.D
Acq On : 22 May 2020 14:28
Operator : JC/SP
Sample : L2718-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
72-12017

Quant Time: May 23 02:38:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

