

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052220\
 Data File : VX016413.D
 Acq On : 22 May 2020 15:36
 Operator : JC/SP
 Sample : L2743-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS042MW019-200521

Quant Time: May 23 03:22:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	209650	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	346061	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	345238	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	178537	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	120803	43.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.58%	
35) Dibromofluoromethane	5.47	113	102170	46.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.20%	
50) Toluene-d8	8.70	98	426725	50.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.02%	
62) 4-Bromofluorobenzene	11.12	95	173249	53.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.84%	
Target Compounds						
3) Chloromethane	1.32	50	1655	0.642	ug/l	95
4) Vinyl Chloride	1.40	62	862	0.314	ug/l	94
16) Acetone	2.42	43	2845	2.311	ug/l	97
27) cis-1,2-Dichloroethene	4.57	96	5015	1.889	ug/l	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052220\
Data File : VX016413.D
Acq On : 22 May 2020 15:36
Operator : JC/SP
Sample : L2743-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS042MW019-200521

Quant Time: May 23 03:22:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

