

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052220\
 Data File : VX016414.D
 Acq On : 22 May 2020 15:59
 Operator : JC/SP
 Sample : L2743-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS042MW024-200521

Quant Time: May 23 03:29:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	215127	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	348942	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	352560	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	175874	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	123277	43.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.10%	
35) Dibromofluoromethane	5.47	113	104593	47.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.62%	
50) Toluene-d8	8.70	98	432136	50.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.46%	
62) 4-Bromofluorobenzene	11.12	95	172926	53.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.76%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.31	50	1024	0.387	ug/l #	86
4) Vinyl Chloride	1.39	62	1009	0.358	ug/l	96
16) Acetone	2.42	43	3254	2.576	ug/l	99
21) trans-1,2-Dichloroethene	3.14	96	1525	0.612	ug/l #	81
27) cis-1,2-Dichloroethene	4.57	96	5353	1.965	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052220\
Data File : VX016414.D
Acq On : 22 May 2020 15:59
Operator : JC/SP
Sample : L2743-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS042MW024-200521

Quant Time: May 23 03:29:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

