

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052220\
 Data File : VX016415.D
 Acq On : 22 May 2020 16:22
 Operator : JC/SP
 Sample : L2743-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z2-TB3-200521

Quant Time: May 23 03:33:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	213719	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	349945	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	341464	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	173324	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	123479	43.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.82%	
35) Dibromofluoromethane	5.46	113	102562	46.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.52%	
50) Toluene-d8	8.70	98	430459	50.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.78%	
62) 4-Bromofluorobenzene	11.12	95	171043	52.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.28%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052220\
Data File : VX016415.D
Acq On : 22 May 2020 16:22
Operator : JC/SP
Sample : L2743-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z2-TB3-200521

Quant Time: May 23 03:33:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

