

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
 Data File : VX035775.D
 Acq On : 22 May 2023 11:14
 Operator : JC/MD
 Sample : VX0522MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0522MBL01

Quant Time: May 23 06:07:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	142015	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	312864	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	283892	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	115598	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	152366	61.500	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	123.000%#
35) Dibromofluoromethane	5.385	113	107296	50.853	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.700%
50) Toluene-d8	8.647	98	362056	46.449	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.900%
62) 4-Bromofluorobenzene	11.079	95	153520	48.916	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
Data File : VX035775.D
Acq On : 22 May 2023 11:14
Operator : JC/MD
Sample : VX0522MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0522MBL01

Quant Time: May 23 06:07:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

